

purposes. The catheter has an optimal length, specific tapered and curved regions, and a progressively tapering diameter along its length. The catheter is made from a hybrid of soft, flexible hydrophilic and reinforced materials to allow for conformational changes in order to accommodate to the variety of vascular anatomy encountered in clinical angiographic practice.